

PARK 500  
PHILIP MORRIS, U.S.A.

SHIFT

D

NAME

M. Johnson

DATE

1/25/90

P

LAB:

Confidential  
Philip Morris v. ABC

- ① NITRATE CALIB. CHECK OOS  $\approx 1330$ .  
CHANNEL WAS RECALIBRATED AND IS NOW  
IN HAND-HELD.

LINE 2: ① L-2 F.S. OOC HIGH FOR DAP. M.R.  
SUPV. WAS NOTIFIED.

- ② 12 REJECTS TOTAL FOR VARIOUS REASONS (LOW PG, STD DEV,  
HIGH OV, LG PICES)

- ③ OFF-STD FOR FIRST 2 PREPS OF THE  
SHIFT (PAGE-4). COLUMN DOWN FOR MAINTENANCE.

LINE 1: ① 2 REJECT FOR OVERT.

LINE 3: ① 4 REJECTS (OV, FELT WASH, IRREGULAR CUT)

2030749305